

The enclosed ampule when properly diluted will give known concentrations for SOME OR ALL of the compounds.

The sample is prepared using the following instructions:

1. Stabilize the ampule to 20°C.
2. Add 1000mL of laboratory grade water to a two (2) liter separatory funnel.
3. Open the ampule, withdraw and add exactly 1.0 mL of the concentrate to the separatory funnel.
4. Mix well.
5. Analyze the resulting solution as if it were a normal sample.
6. Transfer the remaining concentrate to a clean, dry flask or vial and seal. Do not use plastic.

Analyte	Units	WP
2,3,4,6-Tetrachlorophenol	ug/L	50.0 - 200
2,4,5-Trichlorophenol	ug/L	50.0 - 200
2,4,6-Trichlorophenol	ug/L	50.0 - 200
2,4-Dichlorophenol	ug/L	40.0 - 190
2,4-Dimethylphenol	ug/L	65.0 - 200
2,4-Dinitrophenol	ug/L	100 - 180
2,6-Dichlorophenol	ug/L	40.0 - 190
2-Chlorophenol	ug/L	30.0 - 200
2-Methyl-4,6-Dinitrophenol	ug/L	60.0 - 200
2-Methylphenol	ug/L	50.0 - 200
2-Nitrophenol	ug/L	50.0 - 190
3-Methylphenol	ug/L	50.0 - 200
4-Chloro-3-methylphenol	ug/L	30.0 - 200
4-Methylphenol	ug/L	50.0 - 200
4-Nitrophenol	ug/L	100 - 180
Benzoic Acid	ug/L	50.0 - 200
Pentachlorophenol	ug/L	55.0 - 200
Phenol	ug/L	100 - 200



Analytical Products Group, Inc.

2730 Washington Blvd. Belpre, Ohio 45714

Phone 740.423.4200 Fax 740.423.5588

www.apgqa.com www.dmrqa.com