

The sample is prepared from the ampule using the directions below:

1. Stabilize the ampule to 20°C.
2. Add approximately 800 mL of laboratory grade water to a one (1) Liter volumetric flask.
3. Open the ampule, withdraw and add exactly 1.0 mL of the concentrate to the volumetric flask.
4. Dilute to the mark with laboratory grade water.
5. Mix the sample.
6. Analyze the resulting solution as if it were a normal sample.
7. Transfer the remaining concentrate to a clean, dry flask or vial and seal.

Analyte	Units	WS
Endothall	ug/L	90.0 - 500

