

The sample is prepared from the ampule using the directions below:

1. Stabilize the ampule to 20°C.
2. Add 1000 mL of laboratory grade water to a two (2) Liter separatory funnel.
3. Open the ampule, withdraw and add exactly 1.0 mL of the concentrate to the separatory funnel.
4. Mix the sample.
5. Analyze the resulting solution as if it were a normal sample.
6. Transfer the remaining concentrate to a clean, dry flask or vial and seal. Do not use plastic.

Analyte	Units	WS
Aldrin	ug/L	0.400 - 2.00
Dieldrin	ug/L	0.500 - 3.00
Endrin	ug/L	0.100 - 5.00
Heptachlor	ug/L	0.400 - 5.00
Heptachlor epoxide	ug/L	0.200 - 5.00
Hexachlorobenzene	ug/L	0.500 - 4.00
Hexachlorocyclopentadiene	ug/L	2.00 - 30.0
Lindane	ug/L	0.200 - 5.00
Methoxychlor	ug/L	10.0 - 100
Propachlor	ug/L	1.00 - 4.00
Trifluralin	ug/L	1.00 - 5.00

