

Store at 4 °C.

The sample is prepared from the ampule using the directions below:

1. Stabilize the ampule to 20°C.
2. Add approximately 800 mL of laboratory grade water to a one (1) Liter volumetric flask.
3. Open the ampule, withdraw and add exactly 1.0 mL of the concentrate to the volumetric flask.
4. Dilute to the mark with laboratory grade water.
5. Mix the sample.
6. Analyze the resulting solution as if it were a normal sample.
7. Transfer the remaining concentrate to a clean, dry flask or vial and seal.

Analyte	Units	WS
2,4,5-T	ug/L	10.0 - 100
2,4,5-TP (silvex)	ug/L	5.00 - 150
2,4-D	ug/L	5.00 - 150
2,4-DB	ug/L	15.0 - 100
3,5-Dichlorobenzoic acid	ug/L	10.0 - 100
4-Nitrophenol	ug/L	5.00 - 150
Acifluorfen	ug/L	15.0 - 50.0
Bentazon	ug/L	10.0 - 140
Chloramben	ug/L	20.0 - 100
Dacthal diacid (DCPA)	ug/L	20.0 - 100
Dalapon	ug/L	10.0 - 150
Dicamba	ug/L	5.00 - 100
Dichloroprop	ug/L	10.0 - 100
Dinoseb	ug/L	6.00 - 50.0
Pentachlorophenol	ug/L	1.00 - 100
Picloram	ug/L	10.0 - 70.0



Analytical Products Group, Inc.

2730 Washington Blvd. Belpre, Ohio 45714

Phone 740.423.4200 Fax 740.423.5588

www.apgqa.com www.dmrqa.com