

Store at -4°C

The sample is prepared from the ampule using the directions below:

1. Stabilize the ampule to 20°C.
2. Add approximately 800 mL of laboratory grade water to a one (1) Liter volumetric flask.
3. Open the ampule, withdraw and add exactly 1.0 mL of the concentrate to the volumetric flask.
4. Dilute to the mark with laboratory grade water.
5. Mix the sample.
6. Analyze the resulting solution as if it were a normal sample.
7. Transfer the remaining concentrate to a clean, dry flask or vial and seal. Do not use plastic.

Analyte	Units	WS
3-Hydroxycarbofuran	ug/L	15.0 - 75.0
Aldicarb	ug/L	15.0 - 50.0
Aldicarb sulfone	ug/L	19.0 - 50.0
Aldicarb sulfoxide	ug/L	15.0 - 50.0
Baygon (Propoxur)	ug/L	30.0 - 140
Carbaryl	ug/L	20.0 - 100
Carbofuran	ug/L	15.0 - 150
Methiocarb	ug/L	30.0 - 140
Methomyl	ug/L	15.0 - 90.0
Oxamyl (vydate)	ug/L	30.0 - 80.0

