

The enclosed vial when properly diluted will give a sample of known concentration.

To prepare the sample, pipet 10.0 mL of the vial into a one (1) Liter volumetric flask and dilute to the mark with laboratory grade water.

Analyte	Units	WS
MBAS	mg/L	0.0500 - 1.00

