

In order to prepare the sample, add two grams of Sodium Hydroxide and 900 ml of laboratory grade water to a one (1) Liter volumetric flask; then pipet 10 ml of the sample and dilute to the mark with laboratory grade water. Sample distillation is required.

Analyte	Units	DMRQA
Total Cyanide	mg/L	0.100 - 1.00

