

The enclosed vial when properly diluted will give a known concentration.

Mix the vial well before dilution. Pipet 5.0 mL of sample into a one (1) Liter volumetric flask and dilute to the mark with laboratory grade water.

Analyte	Units	Concentration Range
Turbidity	NTU	1.00 - 20.0

